

Examination Glove Nitrile Profi



Description	Tear-resistant Nitrile-Examination glove Length: 30 cm AQL: 1.5, non-sterile Wall Thickness: min. 0,18 mm ~ 7 mil Weight Size M: 11,0 - 12,0 g/piece
Material	Nitrile
Application areas	Technical, cleaning, light chemical protection, medicine, agriculture, especially suitable as milking glove
Product benefits	Protection of the wrist due to extra-long cuff ambidextrous food safe latex- and powderfree free from allergenic thiurams, thiazoles and latex proteins
Single packaging	50 pieces per dispenser box
Standards	KAT III, EN ISO 374-1:2016, EN ISO 374-5:2016 (EU-Regulation 2016/425) EN 455/1-4 (EU-Regulation 2017/745 Class 1 medical device)

Logistical Data:

Item no.	Size	Barcode	Minimum order quantity	Packaging Unit	Pal
15380	S / 6,5-7	4018653153802	1 Pack	1/20	300
15381	M / 7,5-8	4018653153819	1 Pack	1/20	300
15382	L / 8,5-9	4018653153826	1 Pack	1/20	300
15383	XL / 9,5-10	4018653153833	1 Pack	1/20	300
15384	XXL / 10,5-11	4018653153840	1 Pack	1/20	300

Certification:

A) Mechanical resistance according to EN388:2016

Characteristic	Performance level	Range
Abrasion resistance	2	Min 0 / Max 4
Blade cut resistance (Coupe-Test)	0	Min 0 / Max 5
Tear resistance	0	Min 0 / Max 4
Puncture resistance	0	Min 0 / Max 4
Blade cut resistance (according ISO 13997)	X	A-F (X=not tested)

B) Chemical resistance according to EN ISO 374-1:2016, Type A, EN 374-4:2019

ID Code	Substance	CAS-Nr.	Class	Performance level*	Degradation
J	n-heptane	142-82-5	Aliphatic hydrocarbon	6	27,3%
K	40% Sodium Hydroxide	1310-73-2	Inorganic base	6	-11,5 %
M	65 % Nitric Acid	7697-37-2	Inorganic base, oxidising	1	97,8 %
O	25% Ammonium Hydroxide	1336-21-6	Organic base	2	18,8%
P	30% Hydrogen Peroxide	7722-84-1	Peroxide	3	-11,9 %
S	40% Hydrofluoric Acid	7664-39-3	Inorganic acid	2	X
T	37% Formaldehyde	50-00-0	Aldehyde	6	-3,9 %

*legend:

Performance level	1	2	3	4	5	6
Breakthrough time (min)	>10	>30	>60	>120	>240	>480

C) Microbial resistance according to EN ISO 374-5:2016

Protection against bacteria and fungal spores: pass

Protection against virus: pass

D) Medical device according to EN 455/1-4

SRN:	DE-MF-000007708
Basic UDI-DI:	401865300446